

BRIHANMUMBAI MUNICIPAL CORPORATION
CENTRAL PURCHASE DEPARTMENT

e- PROCUREMENT
566, N. M. JOSHI MARG, BYCULLA, MUMBAI -400 011.

No.CPD/Dy. Ch. Eng./Advt./ 46 /TDR /AE-5 Dated 03.06.2024
Corrigendum X

Sub: Amendment to the referred tender enquiry.

Ref.: Dy.Ch. Eng. / CPD/ 68 / TDR/ A.E-05 of 2023- 2024 read with
Corrigendum I, II, III, IV, V, VI, VII, VIII, IX

The following changes are being incorporated in the above referred Tender Enquiry Document.

SECTION - 8 TECHNICAL SPECIFICATIONS		
T.S. No.	Existing	Read as
15.	HPLC system shall be supplied with complete ready to use kit with Buffers in transparent plastic tanks with inventory management system to view the level of buffers; columns, primers, calibrators & sample vials etc.	HPLC system shall be supplied with complete ready to use kit with Buffers in transparent plastic tanks with inventory management system to view the level of buffers; columns, primers, calibrators & sample vials etc. or should have real time inventory management system for monitoring of all reagents & buffers
21.	It should have a sufficient data hard disk approx. 1TB hard drive and a remote data access feature when connected to LAN or Intranet. The system must have a software for real time viewing of the analysis of the sample.	It should have a sufficient data hard disk approx. 1TB hard drive and a remote data access feature when connected to LAN or Intranet or RS-232 . The system must have a software for real time viewing of the analysis of the sample
24.	System should have 2 x RS 232 bidirectional interface and in-built modern for remote service diagnostics / tele services as well as standard pc ports, USB, Ethernet.	System should have 2 x RS 232 bidirectional interface and in-built modern for remote service diagnostics / tele services as well as standard pc ports, USB, Ethernet, RS-232

All other contents of the Tender enquiry including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.

By Order of the
Municipal Commissioner
Brihanmumbai Municipal Corporation

Sd/- 03.06.2024
Dy. Chief Engineer
Central Purchase Department